



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

Weighing/Load Receiving Element
Digital Electronic
Model: BDP Series
 $n_{\max}$: 2 500
 $e_{\min}$: 0.5 lb
Capacity: 1 000 to 2 500 lb
Platform: (see below)
Accuracy Class: III

Submitted By: Contact Info. Updated November 2022

Rice Lake Weighing Systems
230 West Coleman Street
Rice Lake, WI 54868
Tel: 715-234-9171
Fax: 715-234-6967
Contact: Brad Bowers
Email: bbowers@ricelake.com
Web site: www.ricelake.com

Standard Features and Options

The weighing elements covered by this certificate are identified by XxXBDP-YYK, where X x X is the platform dimensions in feet and YY is the capacity in thousand pounds.

Standard Features:

- Platform size up to 3' x 3'
- Length or width can be increased by 125% not to exceed maximum platform area: 9 sq. ft.
- Stainless and Mild Steel

Options:


- Portable Frame


Load Cell Used:

- Sensortronics Model 65023A (Certificate of Conformance Number 86-044A) or equivalent and compatible NTEP Certified Load cell

Temperature Range: 0 °C to 40 °C (32 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.


Randy Jennings
Chairman, NCWM, Inc.


Judith Cardin
Chairman, National Type Evaluation Program Committee
Issued: July 25, 2003

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



Rice Lake Weighing Systems

Weighing/Load Receiving Element / BDP Series

Application: Platform scale for general weighing purposes to be used with an approved and compatible indicating element.

Identification: The required information is located on an adhesive label located on the scale base adjacent to the J-Box cover.

Sealing: The device can be sealed by threading a wire security seal through two fillister head screws on the junction box under the access cover plate.

Test Conditions: The emphasis of the evaluation was on the device design, operation, compliance with influence factor and marking requirements. For the purpose of this evaluation, a Model 3' x 3' BDP-1K (1000 lb x .5 lb), and Model 3' x 3' BDP-2.5K (2500 lb x 1 lb) weighing elements were submitted interfaced with Rice Lake Model IQ+310A indicators (Certificate of Conformance Number 91-123A3). The devices were tested for accuracy over a temperature range of 0 °C to 40 °C (32 °F to 104 °F). Several increasing/decreasing load, shift and corner tests were conducted periodically during this time. A load of approximately one half capacity was applied to each scale base over 100 000 times. The Model 3' x 3' BDP-2.5K was also submitted in a portability frame. Several increasing/decreasing load tests were conducted then the frame was moved 90° and additional increasing/decreasing load tests were conducted. This test was repeated until this scale was tested facing in all four directions.

Evaluated By: A. McCoy, T. Lucas & W. West (OH)

Type Evaluation Criteria Used: NIST, Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices, 2003. NCWM, Publication 14: Weighing Devices, 2003.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: S. Patoray (NCWM), L. Bernetich (NCWM)