



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

Load Cell

Beam

Model: RL1218A Series

 $n_{\max}$: 5000, Class III / Single Cell

Capacity: 50 kg to 1000 kg (110 lb to 2200 lb)

Accuracy Class: III

Submitted By: Contact Info. Updated November 2022

Rice Lake Weighing Systems

230 West Coleman Street

Rice Lake, WI 54868

Tel: 715-234-9171

Fax: 715-234-6967

Contact: Brad Bowers

Email: bbowers@ricelake.comWeb site: www.ricelake.com**Standard Features and Options**

- Nominal Output: 2 mV/V
- 6-wire Design
- Material: Aluminum
- Minimum dead load: 0 kg / 0 lb
- Load Cell Parameters: *capacity evaluated

Capacity (kg)	Capacity (lb)	Single Cell / Class III $n_{\max}$ 5 000 $v_{\min}$ (kg/lb)
50	110	0.0033 / 0.0072
75	165	0.0050 / 0.011
100*	220	0.0067 / 0.015
150	330	0.0100 / 0.022
200	440	0.0133 / 0.030
250	550	0.0167 / 0.037
300	660	0.0200 / 0.045
500	1100	0.0333 / 0.075
635	1400	0.0420 / 0.093
1000	2200	0.0670 / 0.148

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.

Brett Gurney
Chairman, NCWM, Inc.

James Cassidy
Chairman, National Type Evaluation Program Committee
Issued April 16, 2019

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



Rice Lake Weighing Systems

Load Cell / RL1218A Series

Application: The load cells may be used in Class III, single cell applications consistent with the model designations, number of scale divisions, and parameters specified in this certificate. Load cells of a given accuracy class may be used in applications with lower accuracy class requirements provided the number of scale divisions, the $v_{\min}$ value, and temperature range are suitable for the application. The manufacturer may market the load cell with fewer divisions ($n_{\max}$) and with greater $v_{\min}$ values than those listed on the certificate. However, the load cells must be marked with the appropriate $n_{\max}$ and $v_{\min}$ for which the load cell may be used.

Identification: A pressure sensitive identification label located on the cell, states manufacturer name, model and serial number. Other pertinent information will be specified on the Calibration Certificate accompanying the cell.

Test Conditions: This certificate was amended based on additional testing and information from the manufacturer. One 250 kg load cell was tested at NIST Force Group using dead weights as the reference standard. The data were analyzed for single load cell applications. The cell was tested over a temperature range of -10 °C to 40 °C. The temperature effect on zero was measured and a time dependence (creep) test was performed. The cell design is not susceptible to the effects of barometric pressure, therefore this test was not performed. Tests were run on the cell at each temperature. NCWM Publication 14 selection criteria were used to determine cells tested. The results of the testing permitted the adding of the 635 kg and the 1000 kg capacities

Certificate of Conformance Number 18-133: One 100 kg load cell was tested at NIST using dead weights as the reference standard. The data were analyzed for single load cell applications. The cell was tested over a temperature range of -10 °C to 40 °C. The temperature effect on zero was measured and a time dependence (creep) test was performed. The cell design is not susceptible to the effects of barometric pressure, therefore this test was not performed. Tests were run on the cell at each temperature. NCWM Publication 14 selection criteria were used to determine cells tested.

Evaluated By: K. Chesnutwood (NIST Force Group)

Type Evaluation Criteria Used: NIST, Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices, 2019. NCWM, Publication 14: Weighing Devices, 2019.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: J. Truex (NCWM) 18-133; D. Flocken (NCWM) 18-133A1

Examples of Device:

