



## NATIONAL TYPE EVALUATION PROGRAM

# *Certificate of Conformance*

*for Weighing and Measuring Devices*

**For:**

Load Cell

Double Ended Shear Beam

Model: RL75058, RL75060S &amp; RL75058I

 $n_{\max}$  Multiple Cells, Class III: 5 000 $n_{\max}$  Multiple Cells, Class III L: 10 000

Capacity: 20 000 to 200 000 lb

Accuracy Class: III / III L

**Submitted By: Contact Info. Updated November 2022**

Rice Lake Weighing Systems

230 W. Coleman St.

Rice Lake, WI 45868

Tel: 715-234-9171

Fax: 715-234-6967

Contact: Brad Bowers

Email: [bbowers@ricelake.com](mailto:bbowers@ricelake.com)Web site: [www.ricelake.com](http://www.ricelake.com)**Standard Features and Options**

- Material: Alloy Steel (Model RL75058) and Stainless Steel (Model RL75060S)
- The specific load cell capacities,  $v_{\min}$ , and minimum dead loads are listed on Page 2.

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.

Ronald Hayes  
Chairman, NCWM, Inc.

John Gaccione  
Chairman, National Type Evaluation Program Committee  
Issued: December 12, 2014

**1135 M Street, Suite 110 / Lincoln, Nebraska 68508**

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.

**Rice Lake Weighing Systems**

Load Cell / RL75058, RL75060S &amp; RL75058I

**Application:** The load cells may be used for both Class III and III L scales for multiple cell applications consistent with the model designations, number of scale divisions, and parameters specified in this certificate. Load cells of a given accuracy class may be used in applications with lower accuracy class requirements provided the number of scale divisions, the  $v_{\min}$  values, and temperature range are suitable for the application. The manufacture may market the load cell with fewer divisions ( $n_{\max}$ ) and with larger  $v_{\min}$  values than those listed on the certificate. However, the load cells must be marked with the appropriate  $n_{\max}$  and  $v_{\min}$  for which the load cell may be used.

**Load Cell Capacities and Parameters:**

| Capacity<br>(lb) | Class III<br>$v_{\min}$ (lb) | Class III L<br>$v_{\min}$ (lb) | Minimum Dead Load (lb) |
|------------------|------------------------------|--------------------------------|------------------------|
| 20 000           | 1.6                          | 0.60                           | 1000                   |
| 25 000           | 2.0                          | 0.75                           | 1250                   |
| 30 000           | 2.4                          | 0.90                           | 1500                   |
| 35 000           | 2.8                          | 1.05                           | 1750                   |
| 40 000           | 3.2                          | 1.20                           | 2000                   |
| 50 000*          | 4.0                          | 1.50                           | 2500                   |
| 60 000           | 4.8                          | 1.80                           | 3000                   |
| 75 000           | 6.0                          | 2.25                           | 3750                   |
| 90 000           | 7.2                          | 2.70                           | 4500                   |
| 100 000          | 8.0                          | 3.00                           | 5000                   |
| 125 000          | 10.0                         | 3.75                           | 6250                   |
| 150 000          | 12.0                         | 4.50                           | 7500                   |
| 200 000          | 16.0                         | 6.00                           | 10 000                 |

\*One stainless steel and two alloy steel load cells submitted for evaluation.

**Test Conditions:** This Certificate supersedes Certificate of Conformance number 96-027A2 and was issued without additional testing to recognize the change and addition of the model number (RL75058II) specified on the certificate. Previous test conditions are listed below for reference.

**Certificate of Conformance Number 96-027A2:** This Certificate supersedes Certificate of Conformance Number 96-027A1 and is issued to add a Class III, 5000 division and a Class III L, 10 000 division 35 000 lb capacity cell, for multiple cell applications. Justification is based on NCWM Publication 14 criteria for load cells to be tested and a recalculation of the data originally submitted. No additional testing was deemed necessary. The test conditions from the previous certificates are listed below for reference.

**Certificate of Conformance Number 96-027A1:** This Certificate supersedes Certificate of Conformance Number 96-027 and is issued based on the following tests and on information provided by the manufacturer to add a stainless steel construction model to the series. One 50 000 lb capacity stainless steel load cell was tested at NIST using dead weights as the reference standard. The data were analyzed for multiple load cell applications. The cell was tested over a temperature range of  $-10^{\circ}\text{C}$  to  $40^{\circ}\text{C}$ . Three tests were run on the cell at each temperature. The temperature effect on zero was measured and a time dependence (creep) test was performed. The barometric pressure test was waived due to the insensitivity of the load cell design to changes in barometric pressure. This load cell meets all of the requirements for the following two NTEP classifications: Class III L, multiple-cell application, 10 000 divisions and a  $v_{\min}$  of 1.5 lb; and Class III, multiple-cell application, 5000 divisions and a  $v_{\min}$  of 4.0 lb.

**Certificate of Conformance Number 96-027:** This certificate is issued based upon the following tests and upon information provided by the manufacturer. Two load cells of different capacity than those examined for the initial evaluation were tested. New values were assigned for  $v_{\min}$  to meet with current requirements. Two 50 000 lb capacity load cells were tested using dead weights. The data were analyzed for multiple load cell applications. The cells were tested over a temperature range of  $-10^{\circ}\text{C}$  to  $40^{\circ}\text{C}$ . Three tests were run on each cell at each temperature. The temperature effect on zero was measured and a time dependence (creep) test was performed. The barometric pressure test was waived due to the insensitivity of the load cell design to changes in barometric pressure. The manufacturer's laboratory was used to collect the test data. Representatives from the National Institute of Standards and Technology analyzed the data.



## Rice Lake Weighing Systems

Load Cell / RL75058, RL75060S & RL75058I

**Evaluated By:** NIST Force Group, NIST Office of Weights and Measures, 96-027, 96-027A1

**Type Evaluation Criteria Used:** NIST, Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices, 2011. NCWM, Publication 14: Weighing Devices, 2011.

**Conclusion:** The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

**Information Reviewed By:** D. Ripley (NIST) 96-027; G. Newrock (NIST), L. Sebring (NIST) 96-027A1; J. Truex (NCWM) 96-027A2, 96-027A3

**Example of Device:**

