

NIST HB 133 Polyethylene Equipment Proposal

		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	
		0.010058		0.003	

Average = 0.01010 Average = 0.00283
 Avg. Error = 0.00005 Avg. Error = -0.00018
 SD (in) = 0.000031 SD (in) = 0.000041

	MFR.	ID No.
S leaf	Starrett	18456387
Gauge	Mitutoyo	7326S