


Field Standard Meters




Field Standard Meters – Presentation Goals

- Describe - initial calibration traceability
 - Duty meters
 - Field Standard Meters
- Show - additional traceable verification data for Field Standard Meters
 - Gravimetric
 - Volume Proving
- Compare - OIML vs NIST Handbook 44 Tolerances
- Describe - American Petroleum Institute Recommendations
 - Proving and Master Meter
- Describe - Added benefit: Calibration vs Electronic Diagnostic Verification
- Show - Global Field Standard Meter Examples


All Mass Flow Meters Are Calibrated - Metrological Chain For Calibration



Traceable calibration means...
Property of a measurement result
to be **related to a reference**
through a documented and
unbroken chain of calibrations
(99.5% Confidence Level)



Mass Flow Meters Are Calibrated Using Primary and Secondary Standards



Flow calibration via primary (a) and secondary flow standard (b)

OIML Accuracy Class Compared to NIST Handbook 44 3.37 Mass Flow Meter

Table 1

Class	Type of measuring system
0.3	<ul style="list-style-type: none"> - Measuring systems on pipelines (see 5.7) (With exemption for what is stated for accuracy class 1.0 and 1.5)
0.5	All measuring systems, if not differently stated elsewhere in this table, in particular: <ul style="list-style-type: none"> - fuel dispensers for motor vehicles (other than LPG dispensers) (see 5.1, 5.9, and 5.10) - measuring systems on road tankers for liquids of low viscosity (see 5.2) - measuring systems for the unloading of ships' tanks and rail and road tankers (see 5.3) - measuring systems for milk, beer, and other foaming liquids (see 5.6) - measuring systems for loading ships (see 5.7) - measuring systems for refuelling aircraft (see 5.8)
1.0	<ul style="list-style-type: none"> - Measuring systems for liquefied gases under pressure measured at a temperature equal to or above -10°C (see 5.4) - LPG dispensers for motor vehicles (see 5.5) - Measuring systems: <ul style="list-style-type: none"> • used for liquids whose dynamic viscosity is higher than 1000 mPa·s, or • whose maximum flowrate is not higher than 20 L/h or 20 kg/h
1.5	<ul style="list-style-type: none"> - Measuring systems for liquefied carbon dioxide (see 5.4.9), - Measuring systems (other than LPG dispensers) for liquefied gases under pressure measured at a temperature below -10°C (see 5.4)

A better accuracy for a certain type of measuring system may be specified.

Table T.2.
Accuracy Classes and Tolerances for Mass Flow Meters

Accuracy Class	Application or Commodity Being Measured	Acceptance Tolerance	Maintenance Tolerance	Special Tolerance
0.3	<ul style="list-style-type: none"> - Large capacity motor-fuel dispensers (maximum discharge flow rates greater than 100 L/min or 25 gal/min) - Heated products (other than asphalt) at temperatures greater than 50°C (122°F) - Asphalt at temperatures equal to or below 50°C (122°F) - Loading rack meters - Vehicle-tank meters - Home heating oil - Milk and other food products - All other liquid applications not shown in the table where the minimum delivery is at least 700 kg (1500 lb) 	0.2 %	0.3 %	0.5 %
0.3A	<ul style="list-style-type: none"> - Asphalt at temperatures greater than 50°C (122°F) 	0.3 %	0.3 %	0.5 %
0.5	<ul style="list-style-type: none"> - Small capacity (retail) motor-fuel dispensers - Agri-chemical liquids - All other liquid applications not shown in the table where the minimum delivery is less than 700 kg or 1500 lb 	0.3 %	0.5 %	0.5 %
1.0	<ul style="list-style-type: none"> - Anhydrous ammonia - LP Gas (including vehicle-tank meters) 	0.6 %	1.0 %	1.0 %
2.0	<ul style="list-style-type: none"> - Compressed natural gas as a motor-fuel 	1.5 %	2.0 %	2.0 %
2.5	<ul style="list-style-type: none"> - Cryogenic liquid meters 	1.5 %	2.5 %	2.5 %
	<ul style="list-style-type: none"> - Liquefied compressed gases other than LP Gas 			

(Added 1994) (Amended 1999, 2001, and 2013)

OIML System “A” versus Meter “B” Tolerance

Table 2

Line	Accuracy class			
	0.3	0.5	1.0	1.5
A (*)	0.3 %	0.5 %	1.0 %	1.5 %
B (*)	0.2 %	0.3 %	0.6 %	1.0 %
C (equal to Line A – Line B)	0.1 %	0.2 %	0.4 %	0.5 %

(*) see 2.6 for application of line A or line B.


Line A: Complete Measuring System

Line B: Accuracy of one meter


American Petroleum Institute – Manual on Petroleum Measurement Standards

Manual of Petroleum Measurement Standards Chapter 4—Proving Systems


Open Tank



Small Volume Pipe Prover



Piston Prover (Uni or Bi-Di) Prover



Master Meter Prover



American Petroleum Institute – Master Meter Provers

**Manual of Petroleum
Measurement Standards
Chapter 4—Proving Systems**

Section 5—Master-Meter Provers

Measurement Coordination Department

FIRST EDITION, OCTOBER 1988



- Third Edition, June 2011

- Note: Latest issue is:
- Fourth Edition, June 2016

- **Master meter proving** is used when proving by the direct method can not be accomplished because of meter characteristics, logistics, time, space, safety, *and cost considerations*.
- The terminology for this type of proving is “**Indirect master meter proving**”; according to MPMS Ch.4.5 clause 6.5 ,to establish a master meter factor (and this meter could be used for that duty), **a proving against a conventional prover on site or accredited calibration shall be performed** with a repeatability that results in a demonstrated random uncertainty of ±0.029 % or better, at a 95% confidence level.
- Our procedure for proving of the master meter uses an SVP with criteria of 0.05% repeatability in 5 runs : B+B in the table B.2 examples below:

Combinations	Random Uncertainty (RU1) Value of First Meter	Random Uncertainty (RU2) Value of Second	MMF Range (spread) Defined in 6.2 and 6.3	Combined Uncertainty MMF Uncertainty = $\sqrt{RU_1^2 + RU_2^2} + (0.5 \times \text{MMF Range})$
A+B	0.011 %	0.027 %	0.05 %	0.054 %
A+C	0.011 %	0.073 %	0.10 %	0.124 %
C+B	0.073 %	0.027 %	0.15 %	0.153 %
B+B	0.027 %	0.027 %	0.05 %	0.063 %

American Petroleum Institute – Operation of Proving Systems

Manual of Petroleum Measurement Standards Chapter 4—Proving Systems

Section 8—Operation of Proving Systems

Measurement Coordination
FIRST EDITION, NOVEMBER 1995

- Note: Latest issue is:
- 2nd edition, September 2013

4.8.2.11 PROVER RECALIBRATION FREQUENCY

- c. The maximum interval indicated below has elapsed.
1. Three years for small volume provers and mobile provers.
 2. Five years for permanently installed pipe provers.
 3. Five years for permanently installed tank provers.
 4. Three months for master meter provers.

4.8.2.10 FREQUENCY OF METER PROVING

The frequency required for proving varies from several times a day to twice a year or even longer depending upon the value of the liquid, cost/benefit to prove, meter proving history, meter system stability, and variations of operating systems.


For large volumes or different liquids, a permanently installed prover is normally used. The meters should be proved whenever the flow rate, temperature, pressure, API Gravity (relative density), or viscosity changes significantly. Normally, time or volume is used to determine when the meter should be proved.

The proving frequency for new systems should start at short intervals and be extended to longer intervals as confidence increases in the system.

Calibration or Proving Supported By Electronic Diagnostic Verification

What is the relation between calibration and verification results?

Calibration rig




Calibration
report

Field Standard Meter



Verification report

Calibration vs Electronic Diagnostic Verification Comparison




Example: A verification task based on the flowmeter measurement error

Germany Approval for Field Standard Meters

- Approval for use as a field standard of use for testing of measuring systems of accuracy class 0.5*.
- Valid for:
 - Low-viscosity mineral oils (fuels, heating oil, each also with biogenic proportions)
 - Ethanol, Biodiesel
 - Aqueous solutions (including water)
- Approval granted by LBME (Landesbetrieb Mess-und Eichwesen Nordrhein-Westfalen) 12 May 2014

* Compliance with the administrative regulation "Gesetzliches Messwesen – General test instructions for measuring instruments for flowing Liquids other than water (GM-P5)



France Approvals for Field Standard Meters

- Approval for use as a field standard of use for testing of measuring systems
- Valid for:
 - Refined hydrocarbons
 - LPG
- Approval granted by LNE (Laboratoire National De Metrologie Et D'Essais) 11 March 2019

Organisme désigné par
le Ministère chargé de l'Industrie

LABORATOIRE
NATIONAL
DE METROLOGIE
ET D'ESSAIS **LNE**

CERTIFICAT D'APPROBATION DE MOYENS D'ESSAIS
N° LNE-33723 rév. 2 du 11 mars 2019
Modifie le certificat 33723-1

Délivré par : Laboratoire national de métrologie et d'essais
En application : Décret n° 2001-387 du 3 mai 2001 modifié, de l'arrêté du 31 décembre 2001 et de la décision n° 09.00.110.003.1 du 21 octobre 2009 relative aux moyens d'étalonnage dans le domaine du mesurage statique et dynamique des liquides.
Délivré à : SOCIETE GENERALE DE METROLOGIE (SGM) - ZA Bastide Blanche Bâtiment E1
FRANCE 13127 VITROLLES
Concernant : Moyen d'essais SGM type MASS PROVER DEPOT.
Caractéristiques : Les caractéristiques du moyen d'essais sont détaillées en annexe du présent certificat.
Valable jusqu'au : Le présent certificat reste valable tant qu'aucune modification ou évolution susceptible de modifier les caractéristiques du moyen d'essais n'est apportée et sous réserve que l'incertitude globale reste inférieure aux critères fixés pour la catégorie d'instruments concernés. Par ailleurs, toute modification des procédures citées dans le présent certificat doit être portée à la connaissance du LNE.

Les principales caractéristiques et conditions d'approbation figurent dans l'annexe ci-jointe qui fait partie intégrante du certificat d'approbation et comprend 10 pages.
Tous les plans, schémas et notices sont déposés au Laboratoire national de métrologie et d'essais sous la référence de dossier P190168 -1.

Etabli le 11 mars 2019

cofrac
CERTIFICATION
DE PRODUITS
ET SERVICES
Accréditation n° 0-2012
Liste des sites accrédités
et parties approuvées sur
www.cofrac.fr

Pour le Directeur Général
Signature numérique
de THOMAS
LOHMATZSCH ID
Date : 2019.03.11
Thomas LOHMATZSCH +01'00'
Responsable du Pôle Certification
Instrumentation et Technologies de
l'Information

Laboratoire national de métrologie et d'essais • Etablissement public à caractère industriel et commercial
Siège social : 1, rue Gaston Boissier - 75724 Paris Cedex 15 • Tél. : 01 40 43 37 00 - Fax : 01 40 43 37 37
info@lne.fr • lne.fr • RCS Paris 313 320 244 - NAF : 7120B - TVA : FR 92 313 320 244

Netherland Uncertainty Declaration for 10" Field Standard Meter

- Conclusion:
 - The expanded uncertainty of the flowmeter for measuring the quantity of mass flow is 0.064% (bidirectional use) and 0.059% (unidirectional use) with a coverage factor $k=2$ (confidence level of 95%).
 - Approval for use as a field standard of use for testing of measuring systems
- NMI Certain BV, 24 March 2020

NMI Nederlandske Metrische Instansje

Declaration

Number NMI-2400577-EB-02
Project number: 2400577
Page 1 of 1

Issued by: NMI Certin B.V.

Applicant: Endress+Hauser Flowtec AG
Kägenstrasse 7
CH-4153 Reinach (BL)
Switzerland

Submitted: A flow meter intended to be used as master meter for measuring the quantity of mass flow rate.
Manufacturer's identification: Endress+Hauser
Type designation: Promass F300

Scope of assessment: The expanded uncertainty of a master flow meter is determined based on the flow calibration on water and oil products.
This review is solely performed based on water calibration certificates issued by Endress+Hauser under ISO 17025 accreditation no. SCS 0052 by Swiss Accreditation Service (SAS) and the oil product calibration certificates issued by SPSE under ISO 17025 accreditation no. 2.1340 by Cofrac.
Only the determination of the mass flow is in the scope of this assessment.

In accordance with: JCGM 100: 2008
"Evaluation of measurement data — Guide to the expression of uncertainty in measurement".

Annexes: The complete declaration consists of the following annexes:
- Annex 1: General information and uncertainty calculation
- Annex 2: Graphs of calibration results
- Annex 3: Water calibration certificates
- Annex 4: Oil products calibration certificates

Conclusion: The expanded uncertainty of the flow meter for the measuring the quantity of mass flow is 0,064 % (bidirectional use) and 0,059 % (unidirectional use) with coverage factor $k=2$ (confidence level of 95%).

Issuing Authority: NMI Certin B.V., 24 March 2020
E.D. Beumer
Senior Approval Expert

NMI Certin B.V.
Thijssenweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability. Reproduction of the complete document only is permitted.

France



LPG Field Standard Meter



Hydrocarbon Field Standard Meter

France



Hydrocarbon Field Standard Meter

Liquefied Petroleum Gas Meter Proving

Meters used to test Liquefied Petroleum Gas truck mounted delivery system.



Slovenia



Turkey



Trailer Mounted Meter Systems



3" Coriolis Turkey



10" Coriolis South Africa



4" Coriolis Pascagoula MS

Cuba – mobile master meter



Mobile Calibration System in Vietnam



Liquefied Natural Gas Master Meter in Germany



Liquefied Natural Gas Master Meter in Germany



Liquefied Petroleum Gas Master Meter in Slovenia.


Denmark Marketing Terminal – Ship Loading

Application:

- 9 Custody Transfer MID 0.5 approved systems with Coriolis 12”
- 1 Coriolis- 12” Master Meter MID approved on portable trailer
- Fluid: Gasoil, Heavy & Mid Bunker Fuel Oil



Denmark Marketing Terminal – 12” Duty Meter and 12” Master Meter



Denmark Marketing Terminal – 12” Duty Meter and “Master Meter



- Calibration certificate for the duty meter for MID 0.5 application
- Certificate of conformity
- Calibration certificate from accredited laboratory using oil calibration rig with bi-directional ball prover
- Yearly verification certificate against accredited water calibration rig

Mass Flow Meter - Master Meter / Field Standard Meter

- Traceable, accurate and repeatable
 - Traceable water calibration and verification of mass meter is transferrable to the field
 - Suitable for mass and volume proving (P or T compensation for volume proving to standard conditions)
- Reliable
 - Faster, easier, efficient proving compared to using scales or volume provers or displacement provers – easy to automate
 - No mechanical wear, no rotating parts
 - Compact, frame mounted for easy transport
 - On-board electronic verification system ensures ongoing performance between traceable re-verifications
- Field Standard Meters are approved for use worldwide to place into service and test metering systems in contractual and legal for trade applications
- Adoption of LPG-15.1 and MFM 15.1 will lead to broader use in the US and in the future possibly Canada.